



The company **Industrial Scientific Corporation**, Pittsborough, Pennsylvania USA, declares that the following new material:

*(La société Industrial Scientific Corporation, Pittsburgh, Pennsylvania USA, atteste que le matériel neuf désigné ci-après:)*

**Apparatus (Appareil) type VENTIS MX4 6-Unit Calibration Station**

**Complies with the requirements of the following European Directives:**

*(est conforme aux exigences des Directives Européennes suivantes:)*

**I) The European Directive LVD 2014/35/EU of 26/02/14: Low Voltage**

*Directive Européenne Basse Tension 2014/35/EU du 26/02/14*

**Harmonized European Standards:** EN 60950, EN 61010  
*(Normes européennes harmonisées):*

**Test report:** E203424-A1-CB-1  
*(Rapport de test)*

**II) The European Directive EMC 2014/30/EU of 26/02/14: Electromagnetic Compatibility**

*Directive Européenne CEM 2014/30/EU du 26/02/14: Compatibilité Electromagnétique*

**Harmonized European Standards:** EN 61326, EN 55011, Class A  
*(Normes européennes harmonisées):*

**Test report(s):** 1810-7664  
*(Rapport de test)*

**III) The European Directive RoHS 2011/65/EU of 08/06/2011: Restricted Hazardous Substances**

*Directive Européenne RoHS 2011/65/EU du 08/06/2011: Substances Dangereuses Restreintes*

**Harmonised applied standards:** EN 50581:2012  
*(Normes harmonisées appliquées)*

On behalf of the manufacturer  
*Pour le fabricant*

**Industrial Scientific Corporation**  
1 Life Way  
Pittsburgh PA, 15205 USA  
Tel +01 412 788 4353  
[www.indsci.com](http://www.indsci.com)

On behalf of the manufacturer representative in EC  
*Pour le représentant du fabricant dans l'UE*

**Industrial Scientific France SAS**  
5 Rue Frédéric Degeorge, CS 80097  
62002 Arras Cedex, France  
Tel +33 (0)1 57 32 92 61

The ATEX Authorized Representative  
*La Personne Autorisée ATEX*

**Thomas C. Henson**  
Vice President, Products  
*(Vice-président des produits)*  
7 August 2019