

### IN HOUSE CALIBRATION WEIGHTS STATEMENT

Adam Equipment hereby declares that following weights listed below are calibrated within the OIML class tolerance permitted.

Location: Adam UK Weight & Balance test room  
 Description : 1g – 500g cylindrical stainless steel weight set  
 Serial number: AE9RK66  
 Calibrated to OIML Class: F1

OIML Class tolerances:	1g	0.1mg
	2g	0.12mg
	5g	0.16mg
	10g	0.2mg
	20g	0.25mg
	50g	0.3mg
	100g	0.5mg
	200g	0.10mg
	500g	0.25mg

The weights were tested at a temperature of 20° C using standards reference weights with a hypothetical density of 7.9 g/cm<sup>3</sup>.

Reference weights used (UKAS or equivalent):

ID	Mass	Class	Tolerance ( mg)	Uncertainty in ±	Calibration certificate no	Date
ZE2-1 set	1g-500g	E2	See OIML R111-1	See cal cert	M53697	01/07/2020

Master weights on 2 year calibration cycle.

Signed by:



Jon Pooley  
 Dated: 01/07/20