



ADAM EQUIPMENT CO. LTD.

Maidstone Road, Kingston, Milton Keynes, MK10 0BD, UK.
 Tel: (01908) 274545 Fax: (01908) 641339
 Intl Tel: +44 1908 274545 Fax: +44 1908 641339
 sales@adamequipment.co.uk
 www.adamequipment.co.uk

Calibration Certificate

Certificate No: 35784	Make/Model: GFK600	Serial No: AE818580
Capacity: 600kg	Class: N.A.	
Division: 0.05kg	e =	Related Equipment:

Date:	19-May-20
Next Due:	May 21
Tested By:	Richard Harris
Cal Cert Type:	Definitive
Test Weight Set:	CA10,20, AF1&2
Location:	Q.C.

REPEATABILITY		Test Weight used = 500kg			
Definitive:	1. 500.00kg	2. 500.00kg	3. 500.00kg	4. 500.00kg	
	5. 500.00kg	6.	7.	8.	

ECCENTRIC LOADING		Weight = 200kg			Position Sketch				
Definitive:	1. 200.00kg	2. 200.00kg	3. 200.05kg	4. 200.00kg	<table border="1"> <tr> <td>4</td> <td>3</td> </tr> <tr> <td>1</td> <td>2</td> </tr> </table>	4	3	1	2
4	3								
1	2								

	Test Weight	Definitive	
		Increasing Load	Decreasing Load
1	ZERO	0.00kg	0.00kg
2	10kg	10.00kg	10.00kg
3	50kg	50.00kg	50.00kg
4	100kg	100.00kg	100.00kg
5	200kg	200.00kg	200.00kg
6	300kg	300.00kg	299.95kg
7	400kg	400.00kg	399.95kg
8	500kg	500.00kg	499.99kg
9	600kg	600.00kg	
10			

Permissible Tolerances	+/-
	0.10kg

Comments:

The above tests have been carried out with the conditions overleaf, using weights traceable to National Standards.

Signed by: 
 Jon Pooley Service Manager

Date: 10-Aug-20

This certificate must not be reproduced other than in full without prior written approval from Adam Equipment Co Ltd