



Certificate of Compliance

Certificate: 1414649 (LR 46462)

Master Contract: 159134

Project: 1578895

Date Issued: 2004/07/27

Issued to: Siemens Milltronics Process
Instruments Inc.
1954 Technology Dr
Peterborough, Ontario K9J 7B1
Canada
Attention: Mr. Enzo DeSimone

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US'



Issued by: Sam Zaffino

Authorized by: Nick Alfano, Operations
Manager

PRODUCTS

CLASS 2252 83 - PROCESS CONTROL EQUIPMENT - Certified to US Standards
CLASS 2252 03 - PROCESS CONTROL EQUIPMENT

Level Monitor, Model: SITRANS Probe LR or SITRANS LR 200, conduit connected, Enclosure Type 4X / 6,
Rated Input / Output: 24 Vdc Nom., 30 Vdc Max. / 4 – 20 mA loop power, -40 to 80 °C, installation category I,
pollution degree 4.

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 1010.1-92 - Safety Requirements for Electrical Equipment for Measurement, Control, and

The 'C' and 'US' indicators adjacent to the CSA Mark signify that the product has been evaluated to the applicable CSA and ANSI/UL Standards, for use in Canada and the U.S., respectively. This 'US' indicator includes products eligible to bear the 'NRTL' indicator. NRTL, i.e. National Recognized Testing Laboratory, is a designation granted by the U.S. Occupational Safety and Health Administration (OSHA) to laboratories which have been recognized to perform certification to U.S. Standards.



Certificate: 1414649 (LR 46462)

Master Contract: 159134

Project: 1578895

Date Issued: 2004/07/27

Laboratory Use, Part 1: General Requirements

CAN/CSA-C22.2 No. 1010.1B-97 - Amendment 2 to CAN/CSA-C22.2 No. 1010.1-92, "Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements"

ANSI/ISA-82.02.01 1999 - Safety Standards for Electrical and Electronic Test, Measuring, Controlling and Related Equipment - General Requirements



Supplement to Certificate of Compliance

Certificate: 1414649

Master Contract: 159134

The products listed, including the latest revision described below, are eligible to be marked in accordance with the referenced Certificate.

Product Certification History

Project	Date	Description
1578895	2004/07/27	Update to CSA report 1414649 to cover alternate construction
1479591	2003/10/15	Update to cover alternate schematics and pcb artwork.

History

1414649 March 17, 2003 Original Certification