

## Certificate of Compliance

Certificate Number: LR 109492-34C

Revision: -

Date Issued: February 10, 1999

Issued to:

Astec Custom Power (Philippines) Inc.

70 San Rafael St., Barrio Kapitolyo

Pasig City, Philippines

Attention: Ms. Ariette Abarra

The products listed below are eligible to bear the CSA Mark shown



Issued by:

S. Zaffino, C.E.T. Toronto, ON Canada

Signature

PRODUCTS

CLASS 5311 03 - POWER SUPPLIES

Component type power supply for use with Electromedical, Information Technology and Electrical Business Equipment; where the suitability of the combination is determined by the Canadian Standards Association.

Component Power Supply Model:

LPT42-M, Rated input: 100-250V ac, 1.6 Amps, 50/60Hz or 140-300Vdc, 1.0 Amps Output Level 3, +5.0Vdc, 5.0 Amps, +12Vdc, 2.5 Amps and -12Vdc, 0.7 Amps, 40 Watts with Convection Cooling and 55 Watts with 30 CFM Forced Air.

## APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 0-M91 - General Requirements - Canadian Electrical Code, Part II

CAN/CSA-C22.2 No. 234-M90 - Safety of Component Power Supplies

CAN/CSA-C22.2 No. 950-95 - Safety of Information Technology Equipment, Including Electrical

Business Equipment

CAN/CSA-C22.2 No. 601.1-M90 - Medical Electrical Equipment Part I: General Requirements for Safety

CAN/CSA-C22.2 No. AM 2 - Amendment 2 to CAN/CSA-C22.2 No. 601.1-M90

## Conditions of Acceptability

The subject units are only single fused on the input, therefore, fusing in both input line and neutral is required in the end product application.



## Supplement to Certificate of Compliance

Certificate Number: LR 109492-34C

Issued to:

Astec Custom Power (Philippines) Inc.

70 San Rafael St., Barrio Kapitolyo Pasig City, Philippines

Attention: Ms. Arlette Abarra

The products listed, including the latest revision described below, are eligible to be marked in accordance with the referenced Certificate.

Issued By:

S. Zaffino, C.E.T.

Toronto, ON Canada

Signature S. 2/15

**Product Certification History** 

Revision	Date	Description
LR109492-34C	February 10, 1999	Original Certification