

Limited Warranty

Each Fluke product is warranted to be free from defects in material and workmanship under normal use and service. The warranty period is one year and begins on the date of shipment. Parts, product repairs and services are warranted for 90 days. This warranty extends only to the original buyer or end user customer of a Fluke authorized reseller, and does not apply to fuses, disposable batteries or to any product which, in Fluke's opinion, has been misused, altered, neglected or damaged by accident or abnormal conditions of operation or handling. Fluke warrants that software will operate substantially in accordance with its functional specifications for 90 days and that it has been properly recorded on nondefective media. Fluke does not warrant that software will be error free or operate without interruption.

Fluke authorized resellers shall extend this warranty on new and unused products to end user customers only but have no authority to extend a greater or different warranty on behalf of Fluke. Warranty support is available if product is purchased through a Fluke authorized sales outlet or Buyer has paid the applicable international price. Fluke reserves the right to invoice Buyer for importation costs of repair/replacement parts when product purchased in one country is submitted for repair in another country.

Fluke's warranty obligation is limited, at Fluke's option, to refund of the purchase price, free of charge repair, or replacement of a defective product which is returned to a Fluke authorized service center within the warranty period.

To obtain warranty service, contact your nearest Fluke authorized service center or send the product, with a description of the difficulty, postage and insurance prepaid (FOB Destination), to the nearest Fluke authorized service center. Fluke assumes no risk for damage in transit. Following warranty repair, the product will be returned to Buyer, transportation prepaid (FOB Destination). If Fluke determines that the failure was caused by misuse, alteration, accident or abnormal condition of operation or handling, Fluke will provide an estimate of repair costs and obtain authorization before commencing the work.

Following repair, the product will be returned to the Buyer transportation prepaid and the Buyer will be billed for the repair and return transportation charges (FOB Shipping Point).

THIS WARRANTY IS BUYER'S SOLE AND EXCLUSIVE REMEDY AND IS IN LIEU OF ALL OTHER WARRANTIES, EXPRESS OR IMPLIED, INCLUDING BUT NOT LIMITED TO ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. FLUKE SHALL NOT BE LIABLE FOR ANY SPECIAL, INDIRECT, INCIDENTAL OR CONSEQUENTIAL DAMAGES OR LOSSES, INCLUDING LOSS OF DATA, WHETHER ARISING FROM BREACH OF WARRANTY OR BASED ON CONTRACT, TORT, RELIANCE OR ANY OTHER THEORY.

Since some countries or states do not allow limitation of the term of an implied warranty, or exclusion or limitation of incidental or consequential damages, the limitations and exclusions of this warranty may not apply to every buyer. If any provision of this Warranty is held invalid or unenforceable by a court of competent jurisdiction, such holding will not affect the validity or enforceability of any other provision.

In Case of Difficulty

For service or calibration, call your nearest authorized Fluke Service Center. A list of service centers is available on the World Wide Web at www.fluke.com or you can call Fluke using any of the phone numbers listed below.

For application or operation assistance or information on Fluke products, call:

800-44FLUKE (800-443-5853) in the USA and Canada
(31 40) 2678200 in Europe
1 206-356-5500 from other countries

Fluke Corporation
P.O. Box 9090
Everett, WA 98206-9090
U.S.A.

Fluke Europe B.V.
P.O. Box 1186
5602 B.D. Eindhoven
The Netherlands


FLUKE®

700-IV Current Shunt

Instruction Sheet

Introduction

The Fluke 700-IV Current Shunt is used with the 741 or 743 Documenting Process Calibrator to measure and source current simultaneously. With the current shunt, you measure 0 to 55 mA through the calibrator's V MEAS input jacks. Read this sheet before you use the current shunt. This sheet contains specifications and input limit information.



Using the Current Shunt

1. Plug the current shunt into the V MEAS input jacks on the 741 or 743 Calibrator as the figure above shows.
2. Select the measure dc voltage function.
3. Press the **Custom Units** softkey.
4. Press ENTER to select **Current Shunt**, which activates scaling and units programmed for the current shunt.
5. Apply the current source to be measured to the input jacks on the current shunt, and read the results in mA on the display.

Safety Specifications

Designed in accordance with IEC-1010-1; ANSI/ISA S82.01:1994; CAN/CSA C22.2 No. 1010.1 300V CAT II Low energy circuits.

PN 602257
August 1996 Rev.1, 4/97
©1996, 1997 Fluke Corporation. All rights reserved. Printed in USA

Specifications

Input current range	0 to 55 mA dc
Conversion factor	1 mA in converts to 10 mV out
Accuracy (Note)	±0.025% of input
Calibration interval	1 year
Input resistance	250Ω nominal
Output resistance	10Ω nominal
Maximum voltage input	30V dc
Note: Specification applies for 18 to 28 °C. The temperature coefficient for –10 to 18 °C and 28 to 50 °C is 0.001%/°C.	

To obtain the absolute specification for measuring current using the 700-IV and a Model 741 or 743 Calibrator, combine the mV measurement accuracy of the calibrator with the accuracy of the current shunt as shown in the following example.

Assumes input of 20 mA (output = 200 mV), between 18°C and 28°C:

1. Compute the current shunt error term at 20 mA input: 0.005 mA.

This is error term A.
2. From the 731/743 Users Manual look up the 1 year 743 specification for 200 mV:

This is 0.025% of reading +0.005% of full scale. Full scale in this case is 1.1V.

The mV error terms are therefore 0.05 mV + 0.055 mV.

3. Convert the mV error terms to scaled mA equivalents by dividing by 10: 0.005 mA + 0.0055 mA.

These are error terms B and C, respectively.

4. Calculate the absolute specification as follows:

$$\begin{aligned} \text{Absolute specification} &= \pm \left(\sqrt{A^2 + B^2 + C^2} \right) \\ &= \pm 0.013 \text{ mA} \end{aligned}$$

Performance Test

If you need to check the integrity of the current shunt, use the current shunt as described in this instruction sheet to measure an accurate source of 10 mA. Verify that the reading is within specification.

If the current shunt does not work properly, determine the problem as follows:


1. With the current shunt disconnected from the calibrator, measure approximately 250Ω across the current shunt input jacks. An open circuit is probably due to a blown fuse (F1).
2. With the current shunt disconnected from the calibrator, measure approximately 10Ω across the current shunt output plugs. An incorrect resistance indicates damage to any of resistors R1 through R4.

Calibration

To adjust the current shunt, apply 10 mA dc to the current shunt input and adjust Potentiometer R2 for exactly 100 mV dc at the current shunt output.

Cleaning

Periodically wipe the case with a damp cloth and mild detergent. Do not use abrasives or solvents.



Schematic Diagram